



The Commonwealth of Massachusetts

Department of State Police

CHARLES D. BAKER
GOVERNOR

KARYN E. POLITO
LIEUTENANT GOVERNOR

THOMAS A. TURCO, III
SECRETARY

COLONEL CHRISTOPHER S. MASON
SUPERINTENDENT

*Crime Laboratory Office
of Alcohol Testing*

124 Abeton Street

Raynard, Massachusetts 01754

Phone (857) 377-3030 Fax (857) 377-3035

PORTABLE BREATH TEST DEVICES (PBTs)

When making a decision to purchase a portable breath test device (PBT), the Office of Alcohol Testing (OAT) will adjust devices that meet the following criteria:

- 1) The PBT is a Department of Transportation (DOT) approved device
- 2) Data stored on the PBT is downloadable
- 3) The PBT will save the date and time
- 4) The PBT is compatible with Dry Gas for accuracy checks
- 5) The PBT will pass and holds the OAT Calibration Process

Devices which have been previously certified at OAT that do not meet the criteria above may continue to be submitted for adjustment purposes.

The following are approved manufacturers of portable breath testing devices:

Draeger, Inc.

7256 S. Sam Houston W.
Parkway, Suite 100
Houston, TX 77085
(866) 385 – 5900
<http://www.draeger.com>

Lifeloc Technologies, Inc.

12441 W. 49th Avenue, Unit 4
Wheat Ridge, CO 80033
(303) 431 – 9500
<http://www.lifeloc.com>

Intoximeter, Inc.

2081 Craig Road
St. Louis, MO 63146
(800) 451 – 8639
<http://www.intox.com>

Updated on 10/21/2020

Excellence In Service Through Quality Forensic Science