

Significant Changes to WSC-CAM IXB (TO-15): *Quality Control Requirements and Performance Standards for the Analysis of Volatile Organic Compounds in Air Samples (TO-15) by Gas Chromatography/Mass Spectrometry (GC/MS) in Support of Response Actions under the Massachusetts Contingency Plan (MCP)*

- Section 1.6.1 previously required the full analyte list in Table IX B-2 to be reported in order to obtain Presumptive Certainty for TO-15 analyses. This section was revised consistent with other CAM protocols in that the full analyte list is now recommended.