

Significant Changes to WSC-CAM-V A (8082A): *Quality Control Requirements and Performance Standards for the Analysis of Polychlorinated Biphenyls (PCBs) by Gas Chromatography (GC) in Support of Response Actions under the Massachusetts Contingency Plan (MCP)*

- Section 1.6 and Table V A-1 were revised to note that peak area is required for the calibration and quantitation of PCBs. Peak height can only be used if significant interference is present, precluding the use of the peak area.