## **Reminder:** Pharmaceutical Grade or Purpose-Built Refrigerators are Required for <u>All</u> Vaccine Storage Units (Primary and Secondary)

The <u>Guidelines for Compliance with State Vaccine Program Requirements</u> state requires all enrolled practices to have pharmaceutical grade or purpose-built refrigerators for <u>all</u> state-supplied refrigerated vaccine storage units in their facility.

Stand-alone freezers that are not pharmaceutical grade will be acceptable as long as they maintain the proper storage temperature.

Although there is no specific description of a pharmaceutical grade refrigerator, some of the characteristics include:

- Internal overhead fans to disperse air throughout the unit, eliminating pockets of cold air
- Adjustable wire shelves to allow better air flow
- Electronic digital thermostat
- Microprocessor
- No storage bins or shelves on door
- Typically, pharmaceutical grade refrigerators have a narrow operating range (varying by less than 2°C)

Listed below are a sample of manufacturers and distributors that you may wish to consider for safe vaccine storage in your practice. Please note that the Vaccine Program cannot endorse or recommend specific products or brands.

Aegis	http://www.aegisfridge.com
American Biotech Supply	www.americanbiotechsupply.com
Fisher Scientific (Isotemp)	www.fishersci.com
Helmer	www.helmerinc.com
Migali Scientific Refrigeration	www.migaliscientific.com

The Vaccine Program has determined that most standard domestic combination refrigerator/freezer units are inappropriate for vaccine storage. In addition, many stand-alone refrigerators that are not pharmacy grade lack the necessary features for the safe storage of vaccines. Providers identified to be storing any vaccine in non-pharmaceutical grade or purpose-built refrigerators will receive follow-up contacts to ensure that all storage units are in compliance. In addition, as part of the Vaccine Program Restitution Policy, providers may be held financially responsible for vaccine loss due to negligence. This could be significantly more than the cost of a pharmaceutical grade or purpose-built refrigerator.

Before purchasing a new storage unit, utilize the checklist below to ensure that the unit meets the Vaccine Program's requirements.

## Massachusetts Department of Public Health Immunization Division | Vaccine Program

305 South Street, Jamaica Plain, MA 02130 P: 617-983-6828 | F: 857-323-8320 | Email: dph-vaccine-management@mass.gov

## **Vaccine Storage Unit Checklist**

**Complete and return** one checklist for every storage unit that will be holding state-supplied vaccines. **Storage Unit:** 

$\square$ Refrigerator	$\square$ Refrigerator		
(Pharmaceutical)	(Purpose-Built)	□Freezer	□Ultra-Cold Freezer*
Site PIN #:	Brand:		Model #:
	Pof	<u>rigerator</u>	
□Internal fans to disr	perse air throughout the unit	<u>∏Microproc</u>	essor**
☐ Adjustable wire shelves to allow air flow		☐ No storage bins or shelves on door	
☐ Electronic digital thermostat		☐ Refrigerator has a narrow operating range (varying	
		by less than 2° C)**	
	<u> </u>	reezer	,
☐Standalone unit, ca	nnot be mix of fridge and freezer (i	if it is a combination	unit, the fridge and freezer must have
individual plug-ins for	wall outlets)		
$\square$ Must be able to sto	re vaccines below -15° C (5° F)		
	<u>Ultra-G</u>	<u>Cold Freezer</u>	
$\square$ Must be able to ma	intain temperatures between -90°	C and -60° C (-130° F	and -76° F)
$\square$ Sites must privately	purchase a DDL to monitor tempe	ratures in Ultra-Colo	l storage units as they will not be
provided by the State			
*DDLs for Ultra-Cold 1	רemperatures using a buffered tem פו	perature probe provi	ide the most accurate measurement of
vaccine temperatures.	. For accurate ultra-cold temperatu	ire monitoring, it is e	ssential to use an air-probe, or a probe
	or ultra-cold temperatures with the		
		vithin the operating ra	nge, please search the model number online
and review the manufact	turer website for unit specifications.		
Storage Unit Electrical	Plug.		
		ector. extension cor	d, battery pack or a ground fault circuit
interrupter (GFCI) outl		,	,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,
	lugged directly into a wall outlet.		
	,		
Recommendation:			
It is recommended that	t storage units are serviced by a ted	chnician every two y	ears at minimum to ensure proper
functioning and trouble	eshooting of potential issues.		
	·		nd outside of the unit, and any other
relevant information to	assist in making a determination o	on the unit.	

Date \_\_\_\_\_

Completed by: